



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 15-Jan-2018

FR01S12042156-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9, avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1123008
Customer Reference Number:	275/PLANER - 26/12/2017
Product Description:	496611P - DEVICE OOG00199SR148
Validation Reference:	S12029202
Quantity Received:	6
Customer Minimum Specification kGy:	20.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G000693

Irradlatlon Data

Date and Time of Irradiation:	14-JAN-2018 10:18
Calculated Minimum Dose kGy:	28.9
Calculated Maximum Dose kGy:	40.7

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540