



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 27-Sep-2018

FR01S12169448-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
EN ISO 13485 Quality System - Medical Devices

PROMEPLA  
Le Copori  
9 avenue ALbert II  
MONACO  
98000  
MONACO

---

### Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1123008
Customer Reference Number:	306/PLANER PLC - 19/09/2018
Product Description:	496611P - DEVICE OOG00199SR148
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	20.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G001771

---

### Irradiation Data

Date and Time of Irradiation:	26-SEP-2018 00:19
Calculated Minimum Dose kGy:	24.1
Calculated Maximum Dose kGy:	36.3

---

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

---

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540