



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 02-Nov-2015

FR01S11495945-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA

Le Copori
 9, avenue ALbert II
 MONACO
 98000
 MONACO

GR056219/ 1

Qty : 3360

Rohs: ?



2017-05-12

LOT : GR056219/1

PN033938/1

CN101568-1

CN101568-1

QS / BOTTLE

Order information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	194/PLANER - 27/10/2015
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	194773
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	E150913

Irradiation Data

Date and Time of Irradiation:	01-NOV-2015 02:28
Calculated Minimum Dose kGy:	27.7
Calculated Maximum Dose kGy:	37.1

Irradiation Release Authorised By Synergy Health plc

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540