



<http://www.synergyhealthplc.com>

## Certificate of Irradiation

Date Issued: 15-Jun-2015

FR01S11403645-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
 EN ISO 9001:2008 Quality Management System  
 EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA  
 Le Copori  
 9, avenue ALbert II  
 MONACO  
 98000  
 MONACO

---

### Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	170/PLANER - 02/06/2015
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	194773
Quantity Received:	3
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	E150549

---

### Irradiation Data

Date and Time of Irradiation:	13-JUN-2015 12:41
Calculated Minimum Dose kGy:	28.8
Calculated Maximum Dose kGy:	39.2

---

Irradiation Release Authorised By Synergy Health plc

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1