



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 07-May-2018

FR01S12101899-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices


PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1123008
Customer Reference Number:	288/PLANER - 25/04/2018
Product Description:	496611P - DEVICE OOG00199SR148
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	20.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G001261

Irradiation Data

Date and Time of Irradiation:	07-MAY-2018 10:15
Calculated Minimum Dose kGy:	23.3
Calculated Maximum Dose kGy:	39.9

GR064009 / 1	PN039994/1
Qty : 3360	CN101568-1
Rohs : ?	CN101568-1
 2019-11-12	QS / BOTTLE
LOT : GR064009/001	

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1