



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 20-Dec-2017

FR01S12029196-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9, avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023707
Customer Reference Number:	331A-17AB-Steris-carto planer - 27/11/2017
Product Description:	ESSAI - PROMEPLA
Quantity Received:	1
Customer Unit Lot/Batch Number:	OOG00199SR148-G000693
Other Process Details:	Customer required doses : 20.0 - 45.0 kGy

Irradiation Data

Date and Time of Irradiation:	18-DEC-2017 00:25
Calculated Minimum Dose kGy:	25.0
Calculated Maximum Dose kGy:	33.8

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540