



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date issued: 27-Apr-2016

FR01S11611689-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA
 Le Coporl
 9, avenue ALbert II
 MONACO
 98000
 MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	212/PLANER - 25/04/2016
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	194773
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	E160415

Irradiation Data

Date and Time of Irradiation:	27-APR-2016 11:39
Calculated Minimum Dose kGy:	27.1
Calculated Maximum Dose kGy:	36.6

Irradiation Release Authorised By Synergy Health plc

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR69 343 082 840