



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Sep-2021

FR01S12608360-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1134451
Customer Reference Number:	448/PLANER - 22/09/2021
Product Description:	496633P - OOG00799SR148
Validation Reference:	S12538885
Quantity Received:	3
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G005727

Irradiation Data

Date and Time of Irradiation:	28-SEP-2021 03:37
Calculated Minimum Dose kGy:	27.2
Calculated Maximum Dose kGy:	42.5

CN200115 BOTTLE ASSY WITH LABEL (BT37 MK2) Supp Ref:OOG00799SR148	Qty:3,360 Each
Bin:QS/1 GRN:GR3270	Purchase Order #PN238
LOT:G005727 EXP:28/02/2024	MTMS PO # PN046271

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE
N° TVA: FR59 343 092 540

Détail de la commande: 448

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: PLANER PLC					
OOG00799SR148		480	10	G005727	R210124
OOG00799SR148		480	10	G005727	R210125
OOG00799SR148		480	10	G005727	R210126
		1440	30		



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 08-Oct-2021

FR01S12610101-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products

EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1134451
Customer Reference Number:	451/PLANER PLC - 29/09/2021
Product Description:	496633P - OOG00799SR148
Validation Reference:	S12538885
Quantity Received:	4
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G005727

Irradiation Data

Date and Time of Irradiation:	04-OCT-2021 17:11
Calculated Minimum Dose kGy:	28.2
Calculated Maximum Dose kGy:	44.0

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Détail de la commande: 451

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
<i>CLIENT: PLANER PLC</i>					
OOG00799SR148		480	10	G005727	R210129
OOG00799SR148		480	10	G005727	R210130
OOG00799SR148		480	10	G005727	R210131
OOG00799SR148		480	10	G005727	R210132
		1920	40		